

CERTIFICATE

OF

CALIBRATION



ENTERPRISE

IRELAND

ISSUED BY NATIONAL METROLOGY LABORATORY

Glasnevin Dublin 9
Ireland

CLIENT:

Pumpwatch Ltd
PO Box 125
Clane
Co Kildare

Certificate No.: L6105/2
File No.: R6/00008F
Job No.: 116975/2
Order No.: A Byrne

Attention: Alan Byrne

Date Recd: 6/12/99

Description: Pumpwatch CFX20 Prover Vessel

Serial No: Man. No. 948T29

Calibration Temperature: 20°C ± 1.0°C

Calibration	Mettler ID 5	Ser. No	1960982
Standards:	Mettler KB50-2		196088
	OIML Class M1 Masses		202,210
	Comark 9001 Thermometer		728655
	Comark Probe		1171

Method: The CFX20 measure was calibrated by determining the volume of water that is required to fill the dry vessel. The prover was filled to above the overflow with distilled water at a nominal temperature of 20°C. The distilled water volume was determined using a gravimetric method. The overflow valve was then opened and the measured valued compared with the determined value. The measured value was then adjusted to be within ± 1mL of the_determined value.

Calibration Date: 6th December 1999

Date of Issue: 9 December 1999

Calibration By: *Dermot Clancy*
Dermot Clancy

Approved By: *P. Hetherington*
Paul Hetherington

Client work - Terms and Conditions - See overleaf

RESULTS:

The vessel was adjusted to be within ± 1 mL of the nominal value of 20 litres.

COMMENTS:

The volume was determined with an uncertainty of $\pm 0.05\%$ and a confidence of 95%.

All measurements made are traceable through standards maintained by the National Metrology Laboratory, Enterprise Ireland to National and International Standards.